

Pipettes, Graduated

Blaubrand® Class AS

- ◆ To DIN 12697
- ◆ Conformity Certified
- ◆ Calibrated to jet
- ◆ Colour coded
- ◆ Manufactured from AR-Glas® soda-lime glass

Available as:

- ◆ **Blaubrand®** where the graduations and inscriptions are in a highly contrasting blue enamel
- ◆ **Blaubrand® ETERNA** where the graduations and inscriptions are in an amber diffusion stain which forms an integral part of the glass surface. This form of marking is particularly suited to pipettes which have to be subjected to aggressive cleaning solutions

Cap. ml	Grad. ml	Colour coded	Calibration to	Tol. ±ml
0.1	0.001	2 x Green	Contain	0.001
0.5	0.01	2 x Yellow	Deliver	0.005
1	0.01	Yellow	Deliver	0.007
1	0.1	Red	Deliver	0.007
2	0.01	2 x White	Deliver	0.01
2	0.02	Black	Deliver	0.01
2	0.1	Green	Deliver	0.01
5	0.05	Red	Deliver	0.03
5	0.1	Blue	Deliver	0.03
10	0.1	Orange	Deliver	0.05
20	0.1	2 x Yellow	Deliver	0.1
25	0.1	White	Deliver	0.1
50	0.5	—	Deliver	0.2

Graduated Pipettes, Blaubrand®

Soda-lime glass. Calibrated to jet. Graduations and inscriptions in highly contrasting blue enamel. Calibrated to deliver (EX) except 0.1 and 0.2ml capacities which are calibrated to contain (In).


	Ref. 277-	Cap. ml	Grad. ml
Class AS			
PM850-05	02	0.1	0.001
PM850-09	04	0.2	0.002
PM850-14	05	0.5	0.01
PM850-17	06	1	0.01
PM850-20	07	1	0.1
PM850-24	08	2†	0.01
PM850-27	09	2	0.02
PM850-30	10	2	0.1
PM850-35	11	5	0.05
PM850-38	12	5	0.1
PM850-45	13	10	0.1
PM850-50	14	20	0.1
PM850-55	15	25	0.1
PM850-60	16	50	0.5

†Not conformity certified.


Graduated Pipettes, Blaubrand® Eterna

As PM850 but graduations and inscriptions in resistant amber stain. Calibrated to deliver (EX).

	Ref. 284-	Cap. ml	Grad. ml
Class AS			
PM854-14	05	0.5	0.01
PM854-17	06	1	0.01
PM854-27	09	2	0.02
PM854-35	11	5	0.05
PM854-45	13	10	0.1
PM854-55	15	25	0.1



Conformity Certified

Indicated by the symbol  shows that the product has been manufactured by Brand under quality control procedures according to Eichordnung the German Federal Weights and Measures Regulations. Provided the instrument is used with the recommended accessories and in accordance with the supplied instructions, Brand certify conformity with annex 12 of the Eichordnung.