

Thermometers

UKAS Certified Thermometers

UKAS Certified BS1704:1985 (ISO1770:1981) Thermometers

White backed. Spirit filled. The certificate issued provides observed correction at 5 points on the thermometer scale. 75mm immersion.

	Schedule	Range, °C	Grad. °C	Length, mm
TG525-30	D/75 UKAS	0 to 100	1.0	305
TG525-35	E/75 UKAS	0 to 160	1.0	305
TG525-40	F/75 UKAS	0 to 250	1.0	305

Certification of Thermometers

Most types of standard thermometer including BS1704, BS593, IP and ASTM can be certified either as:-

- a) A certificate of conformity.
- or b) The manufacturer's test which provides observed corrections at 5 points on the scale.
- or c) A UKAS (BCS/NPL) test which provides observed corrections at 5 points on the scale.

If you require prices for Certified Thermometers please contact our Sales Department for a quotation.

UKAS – United Kingdom Accreditation Service, formerly NAMAS.

Mercury Thermometers

Under the requirements of Annex XVII of EC552/2009 REACH Regulations, mercury thermometers can no longer be sold within the UK or EU.

Uncertified BS1704 thermometers remain available to special order only and subject to deletion without notice - details on request.

